

BK

Notice of Allowability	Application No.	Applicant(s)
	10/700,086	HENDRIX ET AL.
	Examiner	Art Unit
	Michael P. Stafira	2877

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. This communication is responsive to amendment filed 2/21/2006.
2. The allowed claim(s) is/are 23-36.
3. Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) All
 - b) Some*
 - c) None
 1. Certified copies of the priority documents have been received.
 2. Certified copies of the priority documents have been received in Application No. _____.
 3. Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
5. CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) hereto or 2) to Paper No./Mail Date _____.
 - (b) including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.

Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. Notice of References Cited (PTO-892)
2. Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. Information Disclosure Statements (PTO-1449 or PTO/SB/08),
Paper No./Mail Date _____
4. Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. Notice of Informal Patent Application (PTO-152)
6. Interview Summary (PTO-413),
Paper No./Mail Date _____.
7. Examiner's Amendment/Comment
8. Examiner's Statement of Reasons for Allowance
9. Other _____.

DETAILED ACTION

Allowable Subject Matter

1. Claims 23-36 are allowed over the prior art of record.
2. The following is an examiner's statement of reasons for allowance:

Regarding claim 23, the prior art fails to disclose or make obvious a device for optically inspecting and evaluating a sample having a light source for generating a probe beam, said light source include a first lamp having broadband emissions in the visible spectrum and a second lamp having broadband emissions in the ultraviolet spectrum and a laser diode emitting light in the blue visible region of the spectrum; and a detector for monitoring the reflected probe beam and generating output signals as a function of wavelength, and in combination with the other recited limitations of claim 23. Claims 24-28 are allowed by the virtue of dependency on the allowed claim 23.

Regarding claim 29, the prior art fails to disclose or make obvious a device for optically inspecting and evaluating a sample having a first lamp having broadband emissions in the visible spectrum; a second lamp having broadband emissions in the ultraviolet spectrum; a narrowband light source for emitting narrowband output; means for combining the output from the first and second lamps and the narrowband light source to obtain a broadband probe beam, with the wavelength of the narrowband light source being selected to compensate for a reduced output intensity of the emissions from the first and second lamps; a detector for monitoring the reflected probe beam and generating output signals as a function of wavelength, and in combination with the other recited limitations of claim 29. Claims 30-36 are allowed by the virtue of dependency on the allowed claim 29.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael P. Stafira whose telephone number is 571-272-2430. The examiner can normally be reached on 4/10 Schedule Mon.-Thurs..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Gregory Toatley can be reached on 571-272-2800 ext. 77. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Michael P. Stafira
Primary Examiner
Art Unit 2877

July 26, 2006

Application/Control Number: 10/700,086
Art Unit: 2877

Page 4